

Republika ng Pilipinas
KAGAWARAN NG EDUKASYON
Rehiyon XI
SANGAY NG COMPOSTELA VALLEY
Capitol Complex, Cabidanan, Nabunturan, Compostela Valley




TO : **Whom it may concern**
FROM : **Bids and Awards Committee**
SUBJECT : **Bid Bulletin No. 6**
DATE : **November 22, 2018**

Please be informed of the clarifications in the "PROCUREMENT OF BEAM BALANCE WEIGHING SCALE" to wit:

1. Minimum Capacity: 160kg with non-slip platform pad and one year warranty.
2. Weighing scales must be ISO certified.
3. Unit must be accurate, DOH calibrated and must submit certification of calibration.
4. Place of delivery will be at DeEd, Division Office of Compostela Valley.
5. Sample of the unit shall be presented during the Bid Opening.

These clarifications shall form part of the required compliance in the technical specifications.

For your information and guidance.


JOSIE T. BOLOFER, Ed. D.
BAC Chairperson